

Table S1 MRI acquisition parameters

	MR1	MR2	MR3
Manufacturer	Siemens	Philips	Siemens
Model name	Avanto	Ingenia	Trio Tim
Station name	MRC25494	3FCD991	MRC35363
System version	syngo MR B17	R6.0.531	syngo MR B15
Field strength (T)	1.5	3	3
Head coil channels	8	16	8
3D T1w Images			
Voxel size, mm ³	1.0×1.0×1.0	1.0×1.0×1.0	1.0×1.0×1.0
Number of slices	176	192	176
Repetition time (ms)	1900	7.07	1900
Echo time (ms)	3.37	3.19	2.96
Flip angle (°)	15	7	9

MRI, magnetic resonance imaging; T1WI, T1-weighted images.