

Table S1 The scanning parameters of computed tomography

Parameters	0–20 kg	20–60 kg
Scan method	Helical	Helical
Pitch	0.798	0.798
Tube voltage (kV)	80	100
Tube current (mA)	20–25	30–50
CTDIvol (mGy)	1.6–2.0	2.4–4.0
Scan length (mm)	120	200
Scan FOV (cm × cm)	20×20	25×25
Scan thickness (mm)	5	5
Scan space (mm)	0	0
Scan time (s)	0.4	0.5

CTDIvol, computed tomography dose index; FOV, field of view.