

Table S1 Results of subjective image quality analysis in group A

Item	PCIs	Group A				P value
		40 keV	50 keV	60 keV	70 keV	
Image noise	3 [3, 3]	4 [4, 4]	4 [4, 4]	3 [3, 4]	3 [3, 3]	<0.001 [†]
Vessel sharpness	3 [3, 3]	4 [4, 4.75]	4 [4, 4]	3 [3, 4]	3 [3, 3]	<0.001 [†]
Contrast	3 [3, 3]	4 [4, 5]	4 [3.25, 4]	3 [3, 4]	3 [3, 3]	<0.001 [†]
Diagnostic quality	3 [3, 3.75]	4 [4, 5]	4 [4, 4.75]	3 [3, 4]	3 [3, 4]	<0.001 [†]
Overall image quality	3 [3, 3.75]	4 [4, 5]	4 [4, 4.75]	3 [3, 4]	3 [3, 4]	<0.001 [†]

Data are presented as the median [interquartile range] for ordinal variables. Group A: 8 mL of CM with a 1-mL/s flow rate; group B: 40 mL of CM with 4-mL/s flow rate. [†], Kruskal-Wallis test. PCI, polychromatic conventional image; CM, contrast media.